

Continuous Radon Monitor

Model Number: 1028

Serial Number: 206264023

Calibration Date: 03/31/2023

CF: 2.74

Monitor Time: 6/14/2023 17:25

Inspection Company

Billing Information

Site Information

Phone Number:
License Number:

Site & Condition

Wind: NA

Atmospheric Condition: NA

Year Built: NA

Structure Type: NA

Mitigation System: Not Installed

Monitor Location: NA

SqFt: NA

Test Summary

Start Time: 06/10/2023 09:43


Overall Avg: 2.4 pCi/l

End Time: 06/14/2023 12:43

EPA Avg: 2.4 pCi/l

Measurement Interval(hr): 1.0

Exposure Time: 4 Days 3 hrs



***** 06/10/2023 *****

Time	Counts pCi/l	Flags
10:43	1.5	
11:43	1.1	
12:43	1.8	
13:43	3.3	
14:43	1.1	
15:43	1.5	
16:43	2.6	
17:43	3.3	
18:43	2.2	
19:43	1.1	
20:43	0.0	
21:43	1.1	
22:43	0.7	
23:43	1.1	

***** 06/11/2023 *****

Time	Counts pCi/l	Flags
00:43	2.6	
01:43	3.7	
02:43	2.9	
03:43	3.3	
04:43	2.9	
05:43	3.7	
06:43	2.9	
07:43	3.7	
08:43	3.3	
09:43	2.9	
10:43	2.2	
11:43	3.7	
12:43	2.6	
13:43	4.0	
14:43	1.8	
15:43	5.1	
16:43	1.8	
17:43	1.8	
18:43	0.7	
19:43	2.9	
20:43	2.2	
21:43	2.2	
22:43	2.2	
23:43	3.3	

***** 06/12/2023 *****

Time	Counts pCi/l	Flags
00:43	2.2	

***** 06/12/2023 *****

Time	Counts pCi/l	Flags
01:43	2.6	
02:43	3.7	
03:43	1.5	
04:43	2.2	
05:43	2.2	
06:43	2.2	
07:43	2.2	
08:43	3.7	
09:43	1.1	
10:43	3.3	
11:43	2.6	
12:43	2.2	
13:43	1.1	
14:43	2.2	
15:43	1.5	
16:43	4.4	
17:43	3.3	
18:43	2.6	
19:43	4.4	
20:43	2.9	
21:43	4.0	
22:43	1.5	
23:43	5.1	

***** 06/13/2023 *****

Time	Counts pCi/l	Flags
00:43	1.8	
01:43	0.7	
02:43	2.2	
03:43	1.8	
04:43	2.2	
05:43	3.7	
06:43	4.4	
07:43	1.5	
08:43	2.2	
09:43	1.5	
10:43	2.6	
11:43	3.3	
12:43	2.6	
13:43	1.5	
14:43	1.5	
15:43	1.8	
16:43	1.5	
17:43	2.6	
18:43	2.2	

***** 06/13/2023 *****

Time	Counts pCi/l	Flags
19:43	2.6	
20:43	3.3	
21:43	2.2	
22:43	2.9	
23:43	2.9	

***** 06/14/2023 *****

Time	Counts pCi/l	Flags
00:43	4.4	
01:43	1.5	
02:43	1.5	
03:43	2.9	
04:43	1.1	
05:43	2.2	
06:43	1.8	
07:43	2.9	
08:43	1.8	
09:43	3.3	
10:43	1.1	
11:43	2.2	
12:43	1.5	

Error Flags:
M Motion:

Inspector Signature _____

PC Software Version: 2.2.0
Embedded Software Version: 109